

MICHIGAN DEPT OF AGRICULTURE & RURAL DEVELOPMENT  
 LABORATORY DIVISION  
 WEIGHTS AND MEASURES PROGRAM | MOTOR FUEL QUALITY PROGRAM  
 (517) 655 - 8202  
 michigan.gov/wminfo | michigan.gov/mfq

Device Detail

**Insp Date:** 8/18/2021      **Business ID:** 39665  
**Business:** Freeland Bean & Grain Inc  
 1000 E Washington  
 Freeland, MI 48623

**Inspection:** CD003037  
**Store ID:**  
**Phone:** 989-695-9131  
**Inspector:** 158 DAVID CROWLEY  
**Reason:** FIELD AUDIT

<b>Make:</b> Fairbanks	<b>Type:</b> Heavy Capacity Scale	<b>Results:</b> Sealed
<b>Model:</b> IND-R2500-F2	<b>Subtype:</b> 09-5 (25000)	<b>Other:</b>
<b>Serial:</b> 061740100014	<b>Insp Type:</b> Maintenance	<b>Seal #:</b> Audit Trail
<b>Elapsed Time:</b>	<b>Loc:</b>	<b>Units:</b> lb
<b>d:</b> 20	<b>Capacity:</b> 200000	<b>Sections:</b> 5
<b>CLC or Section Capacity:</b> 40 Tons	<b>Class:</b> IIL	<b>Weighing Elements:</b> Electronic Load Cells
<b>Indicating Elements:</b> DWI	<b>NTEP:</b> Yes	<b>Deck Size:</b> 12 X 70
<b>Invoice#:</b>	<b>Recording Elements:</b> Printer	

Type	Name	Inc/Base	Actual	Display	Err	Tol	Res	Notes
Test	Zero Load	0	0	0	0	20	Pass	
Test	Increasing Load	5000	5000	5000	0	20	Pass	Section 1
Test	Increasing Load	4000	9000	9000	0	20	Pass	
Test	Increasing Load	4000	13000	13000	0	40	Pass	
Test	Shift	0	13000	13000	0	40	Pass	Section 2
Test	Shift	0	13000	13000	0	40	Pass	Section 3
Test	Shift	0	13000	13000	0	40	Pass	Section 4
Test	Shift	0	13000	13000	0	40	Pass	Section 5
Test	Range(Shift)	0	13000		0	40	Pass	
Test	Increasing Load	4000	17000	17000	0	40	Pass	Section 1
Test	Increasing Load	4000	21000	21000	0	60	Pass	
Test	Increasing Load	4000	25000	25000	0	60	Pass	
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 2
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 3
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 4
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 5
Test	Range(Shift)	0	25000		0	60	Pass	
Test	Decreasing Load	-12000	13000	13000	0	40	Pass	Section 1
Test	Zero Load	0	0	0	0	20	Pass	
Test	Strain	31060	25000	56060	0	60	Pass	

**Notes:**  
 Indicator sticker shows Grand Rapids Metrology 8/18/21. Audit trail information:

Cal = 8/18/21, #24677  
 Conf = 6/29/07, #52483

  
 Inspector

\_\_\_\_\_  
 Acknowledged Receipt: