

MICHIGAN DEPT OF AGRICULTURE & RURAL DEVELOPMENT  
 LABORATORY & CONSUMER PROTECTION BUREAU  
 WEIGHTS AND MEASURES PROGRAM | MOTOR FUEL QUALITY PROGRAM  
 (517) 655 - 8202  
 michigan.gov/wminfo | michigan.gov/mfq

Device Detail

**Insp Date:** 6/4/2026      **Business ID:** 39665  
**Business:** FREELAND BEAN & GRAIN INC  
 1000 E Washington  
 Freeland, MI 48623

**Inspection:** CD003766  
**Store ID:**  
**Phone:** 989-695-9131  
**Inspector:** 158 DAVID CROWLEY  
**Reason:** FIELD AUDIT

**Make:** Fairbanks  
**Model:** IND-R2500-F2  
**Serial:** 061740100014  
**Elapsed Time:**  
 d: 20  
**CLC or Section Capacity:** 40 Tons  
**Indicating Elements:** DWI  
**Invoice#:**

**Type:** Heavy Capacity Scale  
**Subtype:** 09-5 (25000)  
**Insp Type:** Maintenance  
**Loc:**  
**Capacity:** 200000  
**Class:** III L  
**NTEP:** Yes  
**Recording Elements:** Printer

**Results:** Sealed  
**Other:**  
**Seal #:** Audit Trail  
**Units:** lb  
**Sections:** 5  
**Weighing Elements:** Electronic Load Cells  
**Deck Size:** 12 X 70

Type	Name	Inc/Base	Actual	Display	Err	Tol	Res	Notes
Test	Zero Load	0	0	0	0	20	Pass	
Test	Increasing Load	5000	5000	5000	0	20	Pass	Section 1
Test	Increasing Load	4000	9000	9000	0	20	Pass	
Test	Increasing Load	4000	13000	13000	0	40	Pass	
Test	Shift	0	13000	13000	0	40	Pass	Section 2
Test	Shift	0	13000	13000	0	40	Pass	Section 3
Test	Shift	0	13000	13000	0	40	Pass	Section 4
Test	Shift	0	13000	13000	0	40	Pass	Section 5
Test	Range(Shift)	0	13000		0	40	Pass	
Test	Increasing Load	4000	17000	17000	0	40	Pass	Section 1
Test	Increasing Load	4000	21000	21000	0	60	Pass	
Test	Increasing Load	4000	25000	25000	0	60	Pass	
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	24980	-20	60	Pass	Section 2
Test	Shift	0	25000	24980	-20	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 3
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 4
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 5
Test	Range(Shift)	0	25000		20	60	Pass	
Test	Decreasing Load	-12000	13000	13000	0	40	Pass	Section 1
Test	Zero Load	0	0	0	0	20	Pass	
Test	Strain	30940	25000	55940	0	60	Pass	

**Notes:**  
 Indicator sticker shows Grand Rapids Metrology 3/26. Audit trail information:

Cal = 8/5/25, #24701  
 Conf = 6/29/07, #52483

  
 Inspector

Acknowledged Receipt :